


optional tank



blue: automatic valve option  
 green : hand operated valve option  
 red : temperature option

Type	Pressure	Volume
1	40,000 psi	20 cc
2	30,000 psi	50 cc
3	10,000 psi	100 cc
4	20,000 psi	125 cc
5	15,000 psi	175 cc
6	10,000 psi	250 cc
7	5,000 psi	500 cc
8	1,875 psi	1000 cc
9	30,000 psi	150 cc
10	20,000 psi	250 cc
11	15,000 psi	300 cc
12	10,000 psi	500 cc
13	5,000 psi	1000 cc

If 2 cylinders

Burt Pressure (PSE): Pressure + 10%

N:\FLOXLAB5 TECHNIQUES\PID5 PUMPS\DUAL CYLINDER PUMP\GENERATION 2\BTD P SERIES V2.DWG



23. RUE DU PORT  
 92022 NANTERRE  
 FRANCE  
 www.floXlab.com

DESIGNATION

# SYRINGE PUMP

DESIGNER: **SAD**      DATE : **11/10/22**

APPROVAL: **PAN**      DATE : **11/10/22**

ETUDE:

**13002**

PAGE: